




Product Service

# CERTIFICATE

No. Z2 16 08 96885 001

**Holder of Certificate:** Xylem Analytics Germany GmbHPeringerstr. 10  
85055 Ingolstadt  
GERMANY**Production Facility(ies):** 96885**Certification Mark:****Product:** Application Software**Model(s):** Data logger systems: Winlog.validation, Winlog.med**Parameters:** Data logger system to read and process data for routine control and validation applications in the field of medicine, pharmacy and food**Tested according to:** TÜV SÜD Product Service testing principles based on

- ISO/IEC 25051:2014 (Functionality)
- DIN EN ISO 9241-110:2008 (Usability)
- FDA 21 CFR Part 11

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 028-713084780-000**Valid until:** 2017-08-31**Date,** 2016-08-11

(Siegfried Muckenschnabl)

Page 1 of 1

